

Product Advantage

- Fulfills the highest requirements for quality
- Provides standardized quality assurance, for example when used in System Suitability Tests (SST)
- Traceable to an official standard
- Provides confirmation of the exactness of the measurement process (accuracy and precision of the results)
- Can be used to compare different Cedex and/or Cedex HiRes systems and, thus, demonstrate the direct comparability of results

Purpose	Beads for one-point density calibration
Density	1 x 10 ⁶ per mL
Size of trading unit	10 mL
Manufacturing	GMP
Traceability	yes (Physikalisch-Technische Bundesanstalt, Berlin)
Product No.	7 60 00 00 17

innovatis AG
 Meisenstr. 96
 33607 Bielefeld
 Germany
 Phone: +49 (0)521 2997-300
 Fax: +49 (0)521 2997-285

innovatis Inc.
 301 Lindenwood Drive, Suite 115
 Malvern, PA 19355
 U.S.A.
 Phone: +1-888-283-1564 Toll free
 Fax: +1-888-283-1631 Toll free
 Phone: +1-610-889-7319
 Fax: +1-610-889-7436

info@innovatis.com
 www.innovatis.com



INNOVATIS
 QUALITY THAT COUNTS

Data specification

Cedex⁺

Density Reference Standard Beads (DRSB)



Product Description

Density Reference Standard Beads are used to calibrate and adjust the Cedex and Cedex HiRes systems. A one-point calibration at 1x10⁶ beads per mL is performed. Generally, the average result from five measurements is compared to the actual value listed on the Certificate of Analysis. Calibration frequency depends strongly on throughput, as well as internal calibration and QA requirements for each customer.